

Contribution ID: 42

Type: not specified

Detecting fusion-evaporation reaction products from contrast agents as a range verification technique in proton therapy (student talk)

Sunday, 18 February 2018 11:15 (15 minutes)

Proton therapy is gaining popularity as a tumor irradiation method due to the superior dose distribution offered by heavy charged particles. The subject of range and dose verification has been approached from several angles and fields.

Upon interaction with low-energy protons, certain nuclei undergo fusion-evaporation reactions. The resulting reaction products emit a cascade of characteristic gamma rays as they decay to their ground state. The product of these fusion-evaporation reactions is highly dependent on the energy of the incident proton. This means that the relative intensity of the competing fusion-evaporation channels can be correlated with the energy of the proton beam, and, by extension, its range. By administering an appropriate contrast agent to the tumor being irradiated, we are able to measure the intensity of characteristic prompt gamma rays resulting from fusion-evaporation reactions occurring inside the tumour.

We propose to take advantage of the timing and energy resolutions of a fast scintillator detector to measure beam range and dose administered in proton radiation therapy. Fast scintillator detectors are used frequently in nuclear research due to their good timing resolution and reasonable energy resolution, and have begun to make their way into the medical field.

Our Geant4 simulation results show that this technique would allow for accurate measurement of beam range relative to the position of the tumour within the body, as well as dose administered to the tumour. Additionally, the excited states resulting from the reactions of interest are very short-lived and as such this measurement is taken while the beam is online. This is advantageous because it allows for dose monitoring during the treatment as opposed to afterwards.

The presented approach combines the unique expertise at TRIUMF in the fields of both gamma ray spectroscopy in nuclear physics research and proton therapy in order to strive towards improved cancer radiation therapy.

Summary

Geant4 simulations of fusion-evaporation reactions between a contrast agent and a proton beam show that we can accurately correlate the relative intensities of competing reaction channels to the range of the proton beam relative to the position of the contrast agent. This technique is applicable to range and dose verification in proton therapy.

Primary authors: Dr MUECHER, Dennis (University of Guelph, TRIUMF); Ms KASANDA, Eva (University of Guelph)

Co-authors: Mr OLAIZOLA, Bruno (TRIUMF); Prof. SVENSSON, Carl (University of Guelph); Mrs BURBADGE, Christina (University of Guelph); Dr HOEHR, Cornelia (TRIUMF); Mr HYMERS, Devin (University of Guelph); Dr

HACKMAN, Greg (TRIUMF); Mr TURKO, Joseph (University of Guelph); Dr BILDSTEIN, Vinzenz (University of Guelph)

Presenter: Ms KASANDA, Eva (University of Guelph)

Session Classification: Session #6