



Contribution ID: 79

Type: **Poster (by default)**

Installation and First Results of a 1.1 kW TWT System for the AECR-U Based Ion Source at UMCG-PARTREC

Due to the demand for intense high-charged stable ion beams for medical and nuclear physics a traveling-wave-tube (TWT) based RF generator has been installed and commissioned at the Advanced Electron Cyclotron Resonance Upgrade (AECR-U) ion source at the UMCG-PARTREC facility. The generator comprises 2 x 750W in-phase combining TWT RF generators with an output frequency range of 12.75-14.5 GHz. Beside its capability to provide the plasma heating needed, it is a very convenient scanning device to identify intense and stable ion-beam regimes within the plasma-heating frequency domain. The new rf generator replaces a 14.1 GHz fixed frequency klystron. Scanning of the plasma-heating frequency allows for an increase of the beam current, for example a factor of 2 for $^{129}\text{Xe}^{17+}$ and 67% for $^4\text{He}^{1+}$ beams, this with respect to former 14.1 GHz fixed frequency. Additionally, stable regimes in the frequency spectrum are identified as the ion-beam stability is monitored at every frequency. In this paper we present the setup, the measurements, and discuss the increase in intensity, fluctuations in stability as well as the overall reproducibility of helium, carbon, and xenon ion beams. These results improve the stability and increase the beam intensity at the UMCG-PARTREC facility.

Author Keywords: [ECR Ion Source] [TWT] [Plasma Heating] [AGOR]

Funding Agency

UMCG-PARTREC

Email Address

h.r.kremers@umcg.nl

I have read the Code of Conduct to attend ICIS2023.

Yes

Presenter if not the submitter of this abstract

H.R. Kremers

Primary authors: Prof. GERBERSHAGEN, Alexander (University Medical Center Groningen); KREMERS, Herman (University Medical Center Groningen)

Co-authors: Mr JONES, Brian (University Medical Center Groningen); Prof. BRANDENBURG, Sytze (University Medical Center Groningen)

Presenter: KREMERS, Herman (University Medical Center Groningen)

Session Classification: Tuesday

Track Classification: Key Technologies for Ion Sources